

United States Department of Commerce
National Institute of Standards and Technology

~ Certificate of Measurement Traceability ~
for

***Demonstrated
Areas of Proficiency**

Mass

Echelon I
20 kg - 1 mg
Echelon II
1200 kg - 1 mg
2500 lb - 1 μ lb
Echelon III
1200 kg - 1 mg
2500 lb - 1 μ lb

Volume

Gravimetric, I
25 gal - 1 fl oz
Volume transfer, II
375 gal - 5 gal

Length

Tapes
500 ft - 1 ft
Rigid rules
48 in - 1 in

Oklahoma

a participant in the NIST Weights and Measures
Measurement Assurance Program for the Year(s)

2009

The State standards are traceable to NIST and the laboratory has demonstrated* measurement proficiency through training and interlaboratory comparisons compliant to NIST HB 143:2007 (ISO/IEC 17025:2005).



Carol T. Hockert,
for the National Institute of Standards and Technology

Chief, Weights and Measures Division

Title

January 1, 2009 to December 31, 2009

Effective Date

