



**State of Oklahoma
Department of Central Services
Procurement**

Amendment of Solicitation

Date of Issuance: September 2, 2009
Requisition No. 1050000119

Solicitation No. 1050000031
Amendment No. 1

Hour and date specified for receipt of offers is changed: No Yes, to: 09/09/2009 4:00 PM CST/CDT

Pursuant to OAC 580:15-4-5(c)(5), this document shall serve as official notice of amendment to the Solicitation identified above. Such notice is being provided to all suppliers to which the original solicitation was sent. Suppliers submitting bids or quotations shall acknowledge receipt of this solicitation amendment prior to the hour and date specified in the solicitation as follows:

- (1) Sign and return a copy of this amendment with the solicitation response being submitted; or,
- (2) If the supplier has already submitted a response, this acknowledgement must be signed and returned prior to the solicitation deadline. All amendment acknowledgements submitted separately shall have the solicitation number and bid opening date printed clearly on the front of the envelope.

ISSUED BY and RETURN TO:

U.S. Postal Delivery:

Department of Central Services, Procurement
P.O. Box 53218
Oklahoma City, OK 73152-3218
or

Keith Hicks
Contracting Officer
(405) - 522 - 3790
Phone Number

Personal or Common Carrier Delivery:

Department of Central Services, Central Purchasing
Will Rogers Building
2401 N. Lincoln Blvd., Suite 212
Oklahoma City, OK 73105

keith_hicks@dcs.state.ok.us
E-Mail Address

Description of Amendment:

a. This is to incorporate the following:

Answers to questions regarding this solicitation are as follows:

Q: Item 1.3.5 states "Must include all components necessary for normal operation. Does this mean that I am to supply a fan and the ductwork for this safety cabinet? If so, we would need a considerable amount of information in regards to where on the roof the fan goes, how far it is from the cabinet, how much ductwork to supply, how many elbows are there to be in the ductwork etc. etc.

A: Item 1.3.5 At the top of the Bid Specification page it states **Free standing, self-contained, Class II, Type A2, Biological Safety Cabinet for use in a Biological Safety Level III microbiological laboratory.**

Q: Item 1.4.2 states "Must maintain freezer at or below -15 degrees centigrade. This item does not come with any type of freezer in it. Is this a mistake?

A: Item 1.4.2 This was just general electrical system requirements, which can be ignored, there is no freezer attached

b. All other terms and conditions remain unchanged.

Supplier Company Name (**PRINT**) _____
Date

Authorized Representative Name (**PRINT**) _____
Authorized Representative Signature