



**STATE OF OKLAHOMA
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Kenneth E. Blick, Ph.D.
Chairman

Post Office Box 36307
Oklahoma City, Oklahoma 73136-2307
Phone: (405) 425-2460 Fax: (405) 425-2490
www.bot.ok.gov

J. Robert Blakeburn
State Director of Tests for
Alcohol and Drug Influence

SPECIAL MEETING AT OKLAHOMA CITY, OKLAHOMA
Oklahoma Department of Public Safety Commissioner's Conference Room
3600 Martin Luther King Ave. DPS building, South East entrance

**March 9, 2012
10:00 A.M. CST
AGENDA**

1. Call to Order and Roll Call
 2. Approve Minutes of meeting August, 31 2011.
 3. Report from State Director
 - A. Consideration of State Directors report
 - B. Motion to approve/disapprove State Directors report
 4. Current Business:
 - A. Consideration or discussion of press publications involving Board personnel matters.
 - B. **Discussion and possible action on the employment of Robert Blakeburn as the State Director of Tests for Alcohol and Drug Influence. The Board may convene an executive session pursuant to 25 OS 307 B 1.**
 - C. Consideration and discussion of amendment to OAC 40:10-9-1 to approve Emergency Medical Technicians licensed at the intermediate, advanced and paramedic levels by the Oklahoma Board of Health to withdraw blood for testing. See attached.
 1. Public comment
 2. Motion to approve, disapprove or table the change.
 - D. Consideration and discussion of approval of modification to the Intoxilyzer I-
- Page one Regular Meeting of Board March 9, 2012

8000 to allow removal of a jumper from the circuit board to prevent failure of the instrument in the event the battery fails and preventing the display from scrolling until the start button is pushed.

1. Public comment
 2. Motion to approve, disapprove or table the modification.
5. Consideration and discussion of approval of modification to the Intoxilyzer I-8000 to allow the instrument to display and/or print data about the time, pressure and/or volume of the breath sample(s) delivered by the individual to be tested.
1. Public comment
 2. Motion to approve, disapprove or table the modification.
6. Consideration and discussion of approval of modification to the Intoxilyzer I-8000 to allow the instrument to retain test data after the information is uploaded, rather than deleting it immediately upon upload.
1. Public comment
 2. Motion to approve, disapprove or table the modification.
7. Report from Chairman
- A. Consideration of Chairman's report
 - B. Motion to approve/disapprove Chairman's report
8. Consideration of Statements by the Public
9. Adjournment